

High Dose-Rate Remote Afterloader Brachytherapy Devices Approved for Patient Treatment Using Sources Exceeding 10 Curies

September 01, 2016

The purpose of this document is to provide a list of high dose-rate (HDR) remote afterloader brachytherapy devices that are FDA-approved for patient treatment using sources exceeding 10 curies (Ci). The NRC will not approve the use of HDR remote afterloaders for patient treatments using sources above 10 Ci if not listed below.

The following HDR remote afterloaders are approved for patient treatment using source activities up to 11 Ci:

- *Mick Radio Nuclear GammaMed (12i, 11i)*
- *Varian VariSource 200 (serial numbers beginning with 200 – 450)*
- *Varian VariSource iX (serial numbers beginning with >500)*

The following HDR remote afterloaders are approved for patient treatment using source activities up to 12 Ci:

- *Eckert & Ziegler BEBIG SangiNova S Version (a.1, a.2, b.1, b.2, c, d)*
- *Mick Radio Nuclear GammaMed (12iT)*
- *Nucletron Flexitron (13149A02)*
- *Nucletron MicroSelectron V2 (105.999)*
- *Nucletron MicroSelectron V3 (106.990)*

U.S. Nuclear Regulatory Commission Contact

Lymari Sepulveda (301) 415-5619

Lymari.Sepulveda@nrc.gov