

X-Ray Machine Inspection Report

RH-F-21 (3/07)

FACILITY ID NO:	PAGE
	OF

FACILITY	FACILITY NAME			SPECIFY PRACTICE TYPE	HOSPITAL	
	ADDRESS				MEDICAL	
	CITY	ST	ZIP		DENTAL	
	PHONE ()	FAX ()	CONTACT PERSON		VETERINARY	
INSPECTOR	NAME			PODIATRIC		
	SIGNATURE			CHIROPRACTIC		
	DATE	PHONE ()		OTHER (SPECIFY)		

TOTAL NO. X-RAY TUBES AT THIS LOCATION	NO. TUBES INSPECTED	NO. OF SERIOUS VIOLATIONS DISCOVERED	NO. OF NON-SERIOUS VIOLATIONS
--	---------------------	--------------------------------------	-------------------------------

OK	N/A	N/S	S	GENERAL ADMINISTRATIVE ITEMS / EVALUATION CRITERIA
----	-----	-----	---	--

OK	N/A	N/S	S	GENERAL ADMINISTRATIVE ITEMS / EVALUATION CRITERIA
				1. Form RH-F-12, "Notice to Employees", not posted (12 VAC 5-481-2260C). NON-SERIOUS.
				2. Personnel radiation dose limits exceeded (<i>explain in remarks</i>) (12 VAC 5-481-640, 700 & 710). SERIOUS.
				3. Radiation exposure levels in unrestricted areas are excessive (<i>explain in remarks</i>) (12 VAC 5-481-720). SERIOUS.
				4. Personnel monitoring records inadequate (<i>explain in remarks</i>) (12 VAC 5-481-680F). SERIOUS. Company Name: _____ Account #: _____ Frequency: Monthly/Quarterly/Other: _____
				5. Personnel monitoring is necessary but not properly provided (12 VAC 5-481-760). SERIOUS.
				6. Safety procedures not posted, inadequate, or not available (12 VAC 5-481-1590-A4). NON-SERIOUS.
				7. Gonadal shielding is necessary, but not available (12 VAC 5-481-1590-A6). SERIOUS.
				8. The speed of the film/screen used is not the fastest consistent with the diagnostic objective (12 VAC 5-481-1590-A9A). NON-SERIOUS.
				8. Information and maintenance records inadequate or not available (12 VAC 5-481-1590-A12). NON-SERIOUS.
				10. Registration form not available for review (12 VAC 5-481-290-1). NON-SERIOUS.
				11. Operator's list not posted, inadequate, or not available (12 VAC 5-481-1590-A14). NON-SERIOUS.
				12. Imaging Practices: _____ Manual _____ Automatic _____ CR _____ Digital Are procedures in place to document film processing practices? <input type="checkbox"/> YES <input type="checkbox"/> NO (12 VAC 5-481-1590-A14). NON-SERIOUS.
				13. Technique chart for the x-ray examination procedure not posted, inadequate, or not available (12 VAC 5-481-1590-A3). NON-SERIOUS.
				14. Other/Remarks: