VDH VIRGINIA DEPARTMENT OF HEALTH
Protecting You and Your Environment

OFFICE OF RADIOLOGICAL HEALTH
P.O. Box 2448 • 7th Floor
Richmond, VA 23218
(804) 864-8150

Dental Panoramic/CBCT Radiographic Systems

PAGE ____ OF ____

RH-F-26 01/2020a

New Registration: ☐

FACILITY			MACHINE			
FACILITY ID NO.	TUBE REGISTRATION NO:		MAKE			
NAME			PURPOSE			
INSPECTOR						
NAME			MODEL			
SIGNATURE			SERIAL NO.			
DATE			MAX. kVp	MAX. mA	ROOM NO.	

OK	N/A	N/S	S	EVALUATION CRITERIA
				1. Warning label not present (12 VAC 5-481-1601-1). **NON-SERIOUS.**
				2. HVL at _____ kVp is measured at _____ mm Al. Minimum is _____ mm Al. (12 VAC 5-481-1601-4). Deficiency of ≤ 0.2mm **NON-SERIOUS**, deficiency of > 0.2mm **SERIOUS**.
				3. Multiple tubes are controlled from one control panel but indication of the selected tube is not present at both the ___ panel and the ___ selected tube (12 VAC 5-481-1601-8). One missing, **NON-SERIOUS,** both missing, **SERIOUS.**
				4. The X-ray field at the plane of the image receptor extends beyond any edge of the image receptor (12 VAC 5-481-1621-F2). **SERIOUS.**
				5. Radiographic control does not require constant operator pressure or does not terminate the exposure properly (12 VAC 5-481-1621-1). **SERIOUS.**
				6. Exposure reproducibility: coefficient of variation _____% at a technique setting of _____ (12 VAC 5-481-1621-B). > 10% to < 15% **NON-SERIOUS,** ≥ 15% **SERIOUS.**
				7. Timer reproducibility: coefficient of variation _____% at a technique setting of _____ (12 VAC 5-481-1621-B). > 10% to < 15% **NON-SERIOUS,** ≥ 15% **SERIOUS**
				8. Radiographic control switch not permanently located in an appropriate protected area defined by a reasonable barrier or distance of at least 9'. (12 VAC 5-481-1621-P2 & P3). **SERIOUS.**
				9. Radiographic control not equipped with both visual and audible indication of X-ray production (12 VAC 5-481-1621-K). Either/or: **NON-SERIOUS,** Both: **SERIOUS**
				10. kVp error ± _____% of indicated kVp at _____ kVp (12 VAC 5-481-1621-A4). > 10% to < 15% **NON-SERIOUS,** ≥ 15% **SERIOUS**
				11. Timer error ± _____ % of indicated seconds at _____ sec (12 VAC 5-481-1621-A4). > 10% to < 15% **NON-SERIOUS,** ≥ 15% **SERIOUS.**
				12. Phantom scatter & other remarks:

_____ mR/hr @ Entrance to alcove; _____ mR/hr @ Operator's station; _____ mR/hr @ Adjacent/Occupied room